



America

CERTIFICATE

No. U8 019054 0009 Rev. 00

Holder of Certificate: LAUDA
DR. R. WOBSE GMBH & CO. KG
 Pfarrstraße 41/43
 97922 Lauda-Königshofen
 GERMANY

Certification Mark:



Product: Temperature controller
 Process Temperature Controllers

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 72159502-000

Date, 2020-08-19

(Richard A Grumski)



America

CERTIFICATE

No. U8 019054 0009 Rev. 00

Model(s): Hei-CHILL 5000 NRTL

Brand Name: LAUDA

Tested according to: CSA C22.2 No. 236:2015
UL 1995:2015

Production Facility(ies): 019054

Parameters:

Rated Input Voltage:	208-220 V AC (1 phase)
Rated Frequency:	60 Hz
Rated Input Current:	16 A
Protection Class:	I

Note:
To be installed as instructed in the installation and Safety Manuals