

# UK CA



## Declaration of Conformity

In accordance with UK Government guidance

This declaration is issued under the sole responsibility of the manufacturer,  
Heidolph Instruments GmbH & Co. KG  
Walpersdorfer Straße 12  
91126 Schwabach / Germany

Product: Shaker & Mixer

Model:

Reax top /control	541-1XXXX-XX
Reax 20/4...12	541-200XX-XX
Reax 2	541-21001-XX
Duomax 1030	543-32XXX-XX
Unimax X010	54X-1XXXX-XX
Promax X020	54X-2XXXX-XX
Polymax X040	54X-4XXXX-XX
Titramax 1XXX	544-1XXXX-XX
Vibramax 1XX	544-X1200-XX
Rotamax 120	544-41200-XX
Multi Reax	545-10000-XX
Inkubator 1000	549-90010-XX

Description:

Shaker & Mixer with different working direction

The object of the declaration described above is in conformity with the relevant UK Statutory Instruments (and their amendments):

2008 No. 1597

*The Supply of Machinery (Safety) Regulations*

2008 2016 No. 1091

*The Electromagnetic Compatibility Regulations 2016*

2012 No. 3032

*The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012*

and complies with the following technical standards :

EN ISO 12100:2010, EN 61326-1:2013, EN 61010-1:2010, EN IEC 63000:2018

UK Authorised Representative (for authorities only)

ProductIP ( UK ) Ltd.

8. Northumberland Av.

London WC2N 5BY

Signed for and on behalf of Heidolph Instruments GmbH & Co. KG  
Walpersdorfer Straße 12, 91126 Schwabach / Germany

Schwabach, 01.09.2021

Wolfgang Jaenicke  
Managing Director

Jörg Ziel  
Quality Manager