

UK CA



Declaration of Conformity

In accordance with UK Government guidance

This declaration is issued under the sole responsibility of the manufacturer,
Heidolph Instruments GmbH & Co. KG
Walpersdorfer Straße 12
91126 Schwabach / Germany

Product: Laboratory vacuum pump

Model:

Hei-VAC Valve Control	591-00130-xx
Hei-VAC Valve Tec	591-00160-xx
Hei-VAC Vario Station	591-00142-xx
Hei-VAC Vario Control	591-00141-xx
Hei-VAC Vario Tec	591-00171-xx
Hei-VAC Valve Industrial	591-07210-xx

Description:

Laboratory vacuum pump to work with laboratory evaporator

The object of the declaration described above is in conformity with the relevant
UK Statutory Instruments (and their amendments):

2008 No. 1597

2008 2016 No. 1091

2012 No. 3032

The Supply of Machinery (Safety) Regulations

The Electromagnetic Compatibility Regulations 2016

*The Restriction of the Use of Certain Hazardous Substances in
Electrical and Electronic Equipment Regulations 2012*

and complies with the following technical standards :

EN ISO 12100:2010, EN 61326-1:2013, EN 61010-1:2010, EN IEC 63000:2018

UK Authorised Representative (for authorities only):

ProductIP (UK) Ltd.
8. Northumberland Av.
London WC2N 5BY

Signed for and on behalf of Heidolph Instruments GmbH & Co. KG
Walpersdorfer Straße 12, 91126 Schwabach / Germany

Schwabach, 26.11.2021

Wolfgang Jaenicke
Managing Director

Jörg Ziel
Quality Manager