

CERTIFICATE OF COMPLIANCE

Certificate Number 2016-12-21-E486210
Report Reference E486210-D1000-1/A0/C0-UL
Issue Date 2016-12-21

Issued to: Heidolph Instruments GmbH & Co. KG
Applicant Company: Walpersdorfer Str. 12,
91126 Schwabach Germany

Listed Company: Same as Applicant

**This is to certify that
representative samples of**

Rotary Evaporator Modular System (Equipment for heating of materials)

Hei-VAP Value Series, Hei-VAP Advantage Series, Hei-VAP Precision Series and Heizbad Hei-VAP (Heating bin).

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety:

UL 61010-1, 2nd Edition, 2013-07, CAN/CSA-C22.2 No. 61010-1, 2nd Edition, 2008-10

Additional Standards:

IEC 61010-2-010:2003(ed.2) - Part 2-010; Particular Requirements for laboratory equipment for heating of materials and CAN/CSA-C22.2 No. 61010-2-010-04 - Part 2-010; Particular Requirements for laboratory equipment for heating of material.

Additional Information:

See the UL Online Certifications Directory at www.ul.com/database for additional information.

Only those products bearing the UL Certification Mark should be considered as being covered by UL's Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.


Bruce Mahrenholz, Assistant Chief Engineer, Global Inspection and Field Services, UL LLC
Joseph Hosey, General Manager, Director of Sales – Canada, UNDERWRITERS LABORATORIES OF CANADA INC.

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative www.ul.com/contactus

